

was effective to grow hair, to prevent falling hair, to correct dandruff, and to revitalize the scalp. The article was not effective for such purposes.

Further misbranding, Section 502 (e) (2), the article was fabricated from two or more ingredients, and the carton label failed to bear the common or usual name of each active ingredient.

DISPOSITION: March 15, 1948. Scalp-Eez, Inc., claimant, having failed to file an answer to the libel or otherwise plead, judgment of condemnation was entered and the product was ordered destroyed.

**2385. Misbranding of Autolift Bust Developers. U. S. v. 303 Autolift Bust Developers, etc.** (F. D. C. No. 24466. Sample No. 14113-K.)

LIBEL FILED: March 18, 1948, Northern District of Illinois.

ALLEGED SHIPMENT: On or about January 30, 1948, by the Flexsaw Co., from Port Austin, Mich.

PRODUCT: 303 *Autolift Bust Developers* at Chicago, Ill., together with a number of circulars entitled "Instruction For Using The Autolift." Examination showed that the product consisted of two plastic cups, ribbon and body strap, and a suction pump with rubber tubing.

NATURE OF CHARGE: Misbranding, Section 502 (a), the following statements in the circular were false and misleading, since the article was not effective in developing the busts: "Autolift Bust Developer. This scientifically designed instrument works on Nature's own principle, suction. This action massages the muscular structure of the breast and stimulates the flow of blood to the desired area. The proper use of this developer will well pay for the trouble, in giving a fuller, rounder, firmer bust. Really an investment in beauty and marital happiness. \* \* \* Exercise busts each night before retiring for best results."

DISPOSITION: May 7, 1948. Default decree of condemnation and destruction.

**2386. Misbranding of Burnett's Radio-Active Emanator. U. S. v. 12 Cones \* \* \*** (F. D. C. No. 23710. Sample No. 49567-H.)

LIBEL FILED: September 23, 1947, Southern District of Mississippi.

ALLEGED SHIPMENT: On or about February 1947, by W. H. Burnett, from Kingsland, Ark.

PRODUCT: 12 devices represented as "Burnett's Radio-Active Emanator" at Decatur, Miss., together with a number of accompanying circulars entitled "Nature's Health Restorer" and "Burnett's Radio-Active Emanator A Health Spring in Your Home." The device was an olive drab-colored solid 10-sided pyramid, about 4½ inches wide at its base and standing about 6 inches high. It consisted of a molded concrete block containing a trace of radioactive material, too little to be of any therapeutic significance.

NATURE OF CHARGE: Misbranding, Section 502 (a), the article was represented by the circulars to be effective in the treatment of kidney disorders, diabetes, high blood pressure, stomach troubles, rheumatism, arthritis, asthma, and other kindred troubles, whereas it was useless for such purposes.

DISPOSITION: March 16, 1948. Default decree of condemnation and destruction.

**2387. Misbranding of Chlorogen devices. U. S. v. 5 Devices, etc.** (F. D. C. No. 23866. Sample Nos. 15007-K, 15008-K.)

LIBEL FILED: October 24, 1947, Eastern District of Michigan.

ALLEGED SHIPMENT: On or about September 2 and 11, 1947, by the Chlorogen Co., from Phoenix, Ariz.

PRODUCT: 5 *Chlorogen devices* at Detroit, Mich., together with a number of pamphlets entitled "Chlorogen Therapy" and one set of operating instructions entitled "Chlorogen Chlorine Gas Generating Inhalator," which were shipped with the devices. Examination of the article showed that it was an electrical device for the production of chlorine.

NATURE OF CHARGE: Misbranding, Section 502 (a), certain statements in the labeling of the device were false and misleading, since they represented and suggested that the device when used as directed was effective in the treatment of sinus infections, upper respiratory diseases, rheumatoid (infectious) arthritis and internal diseases secondary to toxicosis from nasal mucous and sinus infections, sore throat, inflamed tonsils, large goiter, migraine headaches, asthma, sinusitis, bronchitis, and common colds. The device when used as