

4088. Misbranding of Radiant Ozone Generator. U. S. v. 1 Device * * *.
(F. D. C. No. 35233. Sample No. 42570-L.)

LIBEL FILED: May 21, 1953, Northern District of California.

ALLEGED SHIPMENT: In December 1948, by Mrs. Ann McClanahan and her husband, from Eldorado Springs, Mo.

PRODUCT: 1 *Radiant Ozone Generator* at Seaside, Calif. The device consisted essentially of a series of tubes, which were similar to neon tubes, with connections for attachment to a source of electric current.

LABEL, IN PART: "Radiant Ozone Generator Patent No. 2328640 2031 Main St., Kansas City, Mo."

NATURE OF CHARGE: Misbranding, Section 502 (f) (1), the labeling of the device failed to bear adequate directions for use in the treatment of arthritis, angina, diphtheria, mumps, whooping cough, bladder and kidney trouble, blood disorders, anemia, cancer, diabetes, and catarrh, which were the conditions and purposes for which the device was intended to be used.

DISPOSITION: October 7, 1953. Default decree of condemnation and destruction.

4089. Misbranding of Radiant Ozone Generator. U. S. v. 1 Device * * *.
(F. D. C. No. 35234. Sample No. 42568-L.)

LIBEL FILED: May 21, 1953, Northern District of California.

ALLEGED SHIPMENT: In December 1948, by Mr. and Mrs. Roma Hartline, from Eldorado Springs, Mo.

PRODUCT: 1 *Radiant Ozone Generator* at Seaside, Calif. The device consisted essentially of a series of tubes, which were similar to neon tubes, with connections for attachment to a source of electric current.

LABEL, IN PART: "Radiant Ozone Generator Patent No. 2328640 2031 Main St., Kansas City, Mo."

NATURE OF CHARGE: Misbranding, Section 502 (f) (1), the labeling of the device failed to bear adequate directions for use in the treatment of inflammation of the kidneys, neuritis, neuralgia, blood clots, colds, diabetes, sinus trouble, headaches, nervousness, arthritis, sciatic rheumatism, asthma, heart trouble, kidney trouble, boils, poison oak, varicose veins, abnormal blood pressure, stomach cancer, stiff joints, breast cancer, appendicitis, chickenpox, colitis, toxic headaches, high blood pressure, enlarged heart, pleurisy, angina pectoris, asthma, pneumonia, sprains, throat trouble, bruise, cataracts, bloat, eczema, wens, broken bones, liver trouble, stomach trouble, gland trouble, bronchial trouble, rundown condition, cancerous growth, catarrh, constipation, watering eyes, pernicious anemia, paralysis, sore throat, piles, and ear ailment, which were the conditions and purposes for which the device was intended to be used.

DISPOSITION: October 7, 1953. Default decree of condemnation and destruction.

4090. Misbranding of Radiant Ozone Generator. U. S. v. 1 Device * * *.
(F. D. C. No. 35235. Sample No. 42569-L.)

LIBEL FILED: May 21, 1953, Northern District of California.

ALLEGED SHIPMENT: In July 1952, by Mr. Roma Hartline, from Eldorado Springs, Mo.

PRODUCT: 1 *Radiant Ozone Generator* at Seaside, Calif. The device consisted essentially of a series of tubes, which were similar to neon tubes, with connections for attachment to a source of electric current.